



CERTIFICATE OF TEST/CALIBRATION/CONFORMITY TYPE 2

SAEP S'NO: 036973		CUSTOMER ORDER NO: PO0110287		VISUAL/FUNCTIONAL INSPECTION							
				EXTERIOR	✓	INTERIOR	✓	ACCESSORIES		✓	
CUSTOMER: SUB AQUA DIVING SERVICES				SERIAL NUMBER: AB021779							
INSTRUMENT TYPE: O2 PORTABLE				PART NUMBER: MO2BGYY01							
CHANNEL NO	SENSOR NO	TYPE OF GAS	PCB CARD NO	RANGE	ZERO GAS	DISPLAY READING	OUTPUT	SPAN GAS		DISPLAY READING	OUTPUT
1	885096	O2		0-100%	N2	00.1		1	20.85% O2	20.9	
2								2	49.85% O2	49.9	
3								3	100% O2	100.1	
4								4			
INSTRUMENT DETAILS				ACCESSORIES				OPTIONS			
				CHARGER		N		VOLTAGE		BATTERY	
SENSOR TYPE:		9212-5AD		MAINS LEAD		N		MOUNTING		PORTABLE	
CELL SERIAL NO:		885096		BATTERY BRACKETS		N		ALARMS			
SOFTWARE REF:				FLOW ADAPTORS		Y		AUDIO			
RANGE:		0-100%		TUBING		Y		VISUAL			
ALARMS SET:				MANUAL		MO2-807-03		RELAY OPERATION			
				PLASTIC CARDS		Y					
ATMOSPHERIC PRESSURE: 1014mB		TEMP: 23 DEG C		DATE: 21/06/12				GAS CERTIFICATE NO(S): 1263948, 1196479, 1370971			
TEST PROCEDURE:		TESTED BY: 		PRINT NAME: K. FURNISS				APPLIANCE NO(S):			

DSV
~~SADS~~/DOC/AN/01