

CALIBRATION CERTIFICATE

Certificate No : SADS/CAL/460
Item Description : O2 Analyser
Location : DSV DDC
Manufacturer : Analox
SADS ID No : DSV/AN/04
Range : 0-100%
Calibration Device : Analox Cal Gas & SGL Reg


Readings Taken:

Gas	Display Reading	Temperature	Atmospheric Pressure	Alarm Tested
Air % Oxygen	20.9	10 deg C	Unknown	N/A
O2 – 100%	99.8	10 deg C	Unknown	N
Remarks: Check and zero prior to each dive Record this and subsequent tests on 6 monthly sheet.				


Date of Calibration : 9th June 2020

Inspectors Name : N.Graham

Signed :



CERTIFIED COPY



Sub Aqua Diving Services UK Ltd
E: enquiries@commercialdivingservices.co.uk
W: www.commercialdivingservices.co.uk