

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/769
ITEM	02 Analyser
ITEM LOCATION	DSV DDC
SADS ID	DSV/DDC/AN/05
RANGE	0-100%

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gas	Display Reading	Temperature	Atmospheric Pressure	Alarm Tested
02-20.9%	20.9	15 Deg C	Unknown	N/A

Comments:**Remarks:**

Check prior to each dive

) Power on


2) Press & Hold 'Cal' (Cal on air) 100% Elec

3) Oxygen reads 20.9 with flow from LP compressor across the sensor.

4) Recommend calibrating at start of any dive contract and renew sensor and battery as required, due to natural depletion of both.

5) Record this and subsequent tests on 6 monthly sheet.

DATE OF INSPECTION	04 th March 2025	NEXT DUE DATE:	03 rd Sept 2025
---------------------------	-----------------------------	-----------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski**TECHNICIAN SIGNATURE:**

Sub-Aqua
Diving Services (UK) Ltd