

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/763
ITEM	DDC Main Lock Depth Gauge
ITEM LOCATION	DSV-DDC-01
SADS ID	DSV-DDC-DG-01
SERIAL NO. / MODEL NO.	80541/6/11
ACCURACY	0.25%
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	32°C
CALIBRATION DEVICE & SN	Druck DPI 104
CALIBRATION DEVICE ID	SADS/T&T/03
MEASUREMENT UNCERTAINTY	0.25%
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gauge Reading MSW	Druck Reading Bar	Error Bar	Druck Reading Bar	Error Bar
0	0	0	0	0
10	1.00	0	0.98	+0.02
20	1.97	+0.03	1.97	-0.03
30	2.97	+0.03	3.02	-0.02
40	4.03	-0.03	4.05	-0.05
50	4.04	-0.04	4.04	-0.04
Comments:				

DATE OF INSPECTION	04 th March 2025	NEXT DUE DATE:	03 rd Sept 2025
---------------------------	-----------------------------	-----------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski

TECHNICIAN SIGNATURE:
