
Submarine Manufacturing & Products Ltd

Blackpool Road, Newton, Preston Lancashire. PR4 3RE



Email: sales@smp-ltd.co.uk Website: www.smp-ltd.co.uk



	<u>Calibr</u>	ation Certi	ficate.		
Certificate Nº:		80541-6 Date: 21/06/12			
Customer:	SUB AQUA DI	VING SERVICES		/ 33/ ==	
Serial Nº:	80541-6-11				
Type:	150.SB.11.A2.MB.PP				
Range:	230FT/70M(SE	AWATER) A	Accuracy class (%): 0.25		
Test Equipment:	DWT №: 10521/		Certified:	31.12.10	
	Accuracy (%):		~/. 0.04		
	National Standards				
Percentage unce	rtainty:	-/. 0.1			
Condition: x	New. A	s Received.	After Calibration		
Actual Pressur	e.	Actual readings			
m.seawater		Rising.	Fa	Falling.	
0		0		0	
10		10.00	1	10.00	
20		19.95	1	19.95	
30		29.90	2	29.90	
40		39.90	3	39.90	
50		49.90		49.90	
60		59.90	5	59.90	
70		70.00		-	
		×			
			:		
	/				
The reported expanded uncerta	ainty is based in standard	uncertainty multiplied by a	coverage factor k=2, providing	g a level of confidence of	
approximately 95%. The uncertainty	ainty evaluation has been car	ried out in accordance with Ul	KAS requirements.		
Signed:	1-1				
orgineu					