

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/764
ITEM	DDC Main Lock Therapeutic Depth Gauge
ITEM LOCATION	DSV-DDC-01
SADS ID	DSV-DDC-DG-02
SERIAL NO. / MODEL NO.	79799-11-12
ACCURACY	0.25%
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	32°C
CALIBRATION DEVICE & SN	Druck DPI 104
CALIBRATION DEVICE ID	SADS/T&T/03
MEASUREMENT UNCERTAINTY	0.25%
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gauge Reading	Druck Reading	Error	Druck Reading	Error
MSW	Bar	Bar	Bar	Bar
0	0	0	0	0
10	1.00	0	0.99	+0.01
20	1.97	+0.03	2.05	-0.05
30	2.98	+0.02	3.02	-0.02
40	4.01	-0.01	3.99	+0.01
50	5.01	-0.01	4.01	-0.01
Comments:				

DATE OF INSPECTION	04 th March 2025	NEXT DUE DATE:	03 rd Sept 2025
---------------------------	-----------------------------	-----------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski**TECHNICIAN SIGNATURE:** 