

SADS/DOC/DG/01

SADS/DSV/DOC/DG/02 DEPTH GAUGE THERAPUTIC



Submarine Manufacturing &amp; Products Ltd

Blackpool Road, Newton, Preston

Lancashire, PR4 3RE

Tel: 01772 687775 Fax: 01772 687774

Email: sales@smp-ltd.co.uk Website: www.smp-ltd.co.uk

## Calibration Certificate.

Certificate N <sup>o</sup> :	79799-11		Date:	21/06/12
Customer:	SUB AQUA DIVING SERVICES			
Serial N <sup>o</sup> :	79799-11-12			
Type:	150.SB.11.A2.MB			
Range:	80FT/24M(SEAWATER)		Accuracy class (%):	0.25
Test Equipment:	DWT N <sup>o</sup> :	10521/240	Certified:	31/12/10
	Accuracy (%):	+/- 0.04		
Traceable to National Standards			Tested in accordance with BS EN 837	
Percentage uncertainty:		+/- 0.1		

Condition: ☒ New. ☐ As Received. ☐ After Calibration

Actual Pressure. m.seawater	Actual readings	
	Rising.	Falling.
0	0	0
2	2.00	2.00
4	4.00	4.00
6	6.00	6.00
8	8.00	8.00
10	10.00	10.00
12	12.00	12.00
14	13.99	13.99
16	15.99	15.99
18	18.00	18.00
20	19.97	19.97
22	21.97	21.97
24	24.00	-

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Signed: \_\_\_\_\_