

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/765
ITEM	DDC Main Lock Depth Gauge (caisson gauge)
ITEM LOCATION	DSV-DDC-01
SADS ID	DSV-DDC-DG-03
SERIAL NO. / MODEL NO.	81968/11/12
ACCURACY	0.25%
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	32°C
CALIBRATION DEVICE & SN	Druck DPI 104
CALIBRATION DEVICE ID	SADS/T&T/03
MEASUREMENT UNCERTAINTY	0.25%
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gauge Reading	Druck Reading	Error	Druck Reading	Error
MSW	Bar	Bar	Bar	Bar
0	0	0	0	0
10	1.02	-0.02	0.97	+0.03
20	1.96	+0.04	1.02	-0.02
30	3.01	-0.01	3.01	-0.01
40	4.01	-0.01	3.94	+0.06
50	4.04	-0.04	4.02	-0.02
Comments:				

DATE OF INSPECTION	04 th March 2025	NEXT DUE DATE:	03 rd Sept 2025
---------------------------	-----------------------------	-----------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski**TECHNICIAN SIGNATURE:** 