

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020


Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/767
ITEM	5" Low Pressure Oxygen Gauge
ITEM LOCATION	DSV-DDC-01
SADS ID	DSV/DDC-SG-01
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	32°C
CALIBRATION DEVICE	Druck DPI 104
CALIBRATION DEVICE ID	SADS/T&T/04
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Applied Value Bar	Druck Reading Bar	Error Bar	Druck Reading Bar	Error Bar
0	0	0	0	0
5	0.53	-0.03	0.49	+0.01
10	1.05	-0.05	1.0	0
15	1.46	+0.04	1.5	0
20	2.03	-0.03	2.02	-0.02
25	2.48	+0.02	2.47	+0.03
Comments:				

DATE OF INSPECTION	04 th March 2025	NEXT DUE DATE:	03 rd Sept 2025
--------------------	-----------------------------	-------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski

TECHNICIAN SIGNATURE: 

Sub Aqua
Diving Services (UK) Ltd