

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/768
ITEM	5" Low Pressure Air Supply Gauge
ITEM LOCATION	DSV-DDC-01
SADS ID	DSV/DDC-SG-02
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	10 Deg C
CALIBRATION DEVICE	Druck DPI 104
CALIBRATION DEVICE ID	SADS/T&T/03
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Applied Value Bar	Druck Reading Bar	Error Bar	Druck Reading Bar	Error Bar
0	0	0	0	0
5	0.53	-0.03	0.52	-0.02
10	1.03	-0.03	1.0	0
15	1.49	+0.01	1.56	-0.06
20	2.03	-0.03	2.05	-0.05
25	2.44	+0.06	2.48	+0.02
Comments:				

DATE OF INSPECTION	04 th March 2025	NEXT DUE DATE:	03 rd Sept 2025
--------------------	-----------------------------	-------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski

TECHNICIAN SIGNATURE: 