SAR-DSV-DOC-SG-02 LP SUPPLY GAUGE

Submarine Manufacturing & Products Ltd Blackpool Road, Newton, Preston Lancashire. PR4 3RE



Tel: 01772 687775 Fax: 01772 687774 Email: sales@smp-ltd.co.uk Website: www.smp-ltd.co.uk

Calibration Certificate.

Certificate Nº:	81014-3						21/0	6/12	
Customer:	81014-3 Date: 21/06/12 SUB AQUA DIVING SERVICES								
Serial Nº:	81014-3-11								
Type:	100.SB.11.A2								
Range:		400 PSI/BAR Accuracy				class (%):	1.0	
Test Equipment:	DWT Nº:		10521/240			Certified: 31.12.10		31.12.10	
Accurac		y (%):				⁴/. 0.04			
	National Standa							S EN 837	
Percentage uncertainty: 7/. 0.25									
Condition: X New. As Received. After Calibration									
Actual Pressure. BAR		Actual readings							
		Rising.				Falling.			
	l l								
0		0			0				
5		5.00				5.00			
10		10.10				10.10			
15		15.10				15.10			
20		20.00				02.00			
25		25.00				25.00			
28		28.00				-			
								1	
			1						
				2,000,000 to 100 to					
The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of									
approximately 95%. The uncertainty-evaluation has been carried out in accordance with UKAS requirements.									
Signed:									