

## **CALIBRATION CERTIFICATE**

Doc Ref: SADS-IMS-FM-90 Issue 01

Issue Date: 28/05/2020 Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/779		
ITEM	02 Analyser		
ITEM LOCATION	DSV Dive Control Panel		
SADS ID	DSV/DP/AN/02		
RANGE	0-100%		
MANUFACTURER	Greisinger		

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gas	Display Reading	Temperature	Atmospheric Pressure	Alarm Tested
02-20.91%	20.9	10 Deg C	Unknown	N/A

## **Comments:**

## **6 Monthly Calibration**

## Remarks:

Check prior to each dive

- ) Power on
- 2) Press & Hold 'Cal' (Cal on air) 100% Elec
- 3) Oxygen reads 20.9 with flow from LP compressor across the sensor.
- 4) Recommend calibrating at start of any dive contract and renew sensor and battery as required, due to natural depletion of both.
- 5) Record this and subsequent tests on 6 monthly sheet.

	A 100 A	where he has been a second	
DATE OF INSPECTION	05 <sup>th</sup> March 2025	NEXT DUE	04 <sup>th</sup> Sept 2025
	The state of the s	DATE:	

TECHNICIAN NAME: Dan Zapalski

TECHNICIAN SIGNATURE:

Tech Dept

Scholingury Stamp

Diving Services

Sub Aqua Diving Services UK Ltd
E: <a href="mailto:enquiries@commercialdivingservices.co.uk">enquiries@commercialdivingservices.co.uk</a>
W: <a href="mailto:www.commercialdivingservices.co.uk">www.commercialdivingservices.co.uk</a>