

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90
Issue 01
Issue Date: 28/05/2020
Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/772
ITEM	Depth Gauge Diver 1
ITEM LOCATION	DCV Dive panel
SADS ID	DSV-DP-DG-01
SERIAL NO. / MODEL NO.	322766-3
ACCURACY	0.25%
CALIBRATION POSITION	Vertical
AMBIENT TEMPRATURE	18°C
CALIBRATION DEVICE & SN	SADS-T&T-03 .3108781. Tested 22 Oct 2024
MEASUREMENT UNCERTAINTY	0.25%
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gauge Reading MSW	Druck Reading Bar	Error Bar	Druck Reading Bar	Error Bar
0	0	0	0	0
5	.494	+ .006	0.494	+ .006
10	.995	+ .005	1.010	- .010
15	1.504	- .004	1.504	- .004
20	2.013	- .013	2.017	- .017
30	3.006	- .006	3.017	- .017
40	4.016	- .016	4.022	- .022
60	6.005	- .005	6.051	- .051
70	7.016	- .016	7.016	- .016
Comments:				

DATE OF INSPECTION	05 th March 2025	NEXT DUE DATE:	04 th Sept 2025
---------------------------	-----------------------------	-----------------------	----------------------------

TECHNICIAN NAME: Dan Zapalski/Derek Curtis

TECHNICIAN SIGNATURE:


