



Ashford Instrumentation Ltd
Units 2 & 3 Fairview Industrial Park
Hamstreet Road
Ruckinge, Nr. Ashford
Kent, TN26 2PL
Tel: 0044 (0)1233 730999
Fax: 0044 (0)1233 731133
Email: sales@ashfordinstrumentation.com

CERTIFICATE OF CALIBRATION

PRODUCT DETAILS

CERT TYPE	Definitive	CERTIFICATE No.	322766/3
PRODUCT	Depth Gauge	DATE	14/01/2013
SERIAL No.	322766-3	ACCURACY +/- FS	0.25%
DIAL RANGE	0 to 70 m & ft of Sea Water @ SG 1.025	ACCURACY UNITS	0.175
TEST RANGE 0/	70	REC. CAL. FREQ'Y	1 Year
TEST UNITS	m sea water	SALES O/No.	322766
SPECIFICATION	EN837	CAL. MEDIUM	Clean Air

RISING RESULTS

FALLING RESULTS

PRODUCT READING	STANDARD READING	CORRECTION ERROR	STANDARD READING	CORRECTION ERROR
m sea water	m sea water	m sea water	m sea water	m sea water
0	0	0.00	0	0.00
5	5	0.00	5	0.00
10	10	0.00	10	0.00
15	15	0.00	15	0.00
20	20	0.00	20	0.00
25	25	0.00	25	0.00
30	30	0.00	29.9	-0.10
35	35	0.00	35	0.00
40	40	0.00	40	0.00
45	45	0.00	45	0.00
50	50	0.00	50	0.00
55	55.1	0.10	55.1	0.10
60	60.1	0.10	60.1	0.10
65	65.1	0.10	65.1	0.10
70	70.1	0.10		

The above product has been tested and found to conform to the specifications, unless stated below.
COMMENTS:-

DETAILS OF CALIBRATION EQUIPMENT USED

STANDARD	Dead Weight Tester	UKAS LAB No.	0134
SERIAL No.	696A A9760	CERT No.	08WA9760
ACCURACY	0.04%		
EXPIRY DATE	12/09/2013		

The calibration equipment used is within its period of calibration and is traceable to NAMAS/UKAS Standards.

Signed

Authorised Signature



Company Registration Details: Company Limited, Incorporated May 2003, Registered in England
Registration Number 475 6319, Registered Office Same As Trading Address

