

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01

Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/773
ITEM	Diver 2 depth gauge
ITEM LOCATION	DSV dive panel
SADS ID	DSV-DP-DG-02
SERIAL NO. / MODEL NO.	322766-2
ACCURACY	0.25%
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	18°C
CALIBRATION DEVICE & SN	SADS-T&T-03 .3108781. Tested 22 Oct 2024
CALIBRATION DEVICE ID	SADS/T&T/03
MEASUREMENT UNCERTAINTY	0.25%
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gauge Reading MSW	Druck Reading Bar	Error Bar	Druck Reading Bar	Error Bar
0	0	0	0	0
5	.506	-.004	.513	-.007
10	.998	+.002	1.024	-.024
15	1.494	+.006	1.532	-.032
20	2.016	-.016	2.052	-.052
30	3.023	-.023	3.056	-.056
40	4.032	-.032	4.052	-.052
60	6.033	-.033	6.086	-.086
70	7.052	-.052	7.053	-.053
Comments:				

DATE OF INSPECTION05th March 2025**NEXT DUE DATE:**04th Sept 2025**TECHNICIAN NAME:** Dan Zapalski/ Derek Curtis**TECHNICIAN SIGNATURE:**