

CALIBRATION CERTIFICATE

Doc Ref: SADS-IMS-FM-90

Issue 01


Issue Date: 28/05/2020

Issued By: P. Curtis

CERTIFICATE NUMBER	SADS/CAL/777
ITEM	Diver 3 supply gauge
ITEM LOCATION	SADS/DSV/DC/01
SADS ID	DSV-DP-SG-03
ACCURACY	0.25%
CALIBRATION POSITION	Vertical
AMBIENT TEMPERATURE	18 degrees C
CALIBRATION DEVICE & SN	SADS-T&T-03 .3108781. Tested 22 Oct 2024
CALIBRATION DEVICE ID	SADS/T&T/03
MEASUREMENT UNCERTAINTY	0.25%
TEST SPECIFICATION	BS EN 837-1:1998

READINGS TAKEN:	RISING RESULTS		FALLING RESULTS	
Gauge Reading	Druck Reading	Error	Druck Reading	Error
PSI	PSI	PSI	PSI	PSI
50	52	-2	50	0
100	105	-5	103	-3
200	206	-6	206	-6
300	305	-5	304	-4
Comments:				

DATE OF INSPECTION	05 th March 2025	NEXT DUE DATE:	04 th Sept 2025
---------------------------	-----------------------------	-----------------------	----------------------------

TECHNICIAN NAME: D Zapalski/D Curtis**TECHNICIAN SIGNATURE:**

Sub Aqua
Diving Services (UK) Ltd