

CERTIFICATE OF CALIBRATION

Eurocentre, North River Road
Great Yarmouth
Norfolk, NR30 1TE
Tel No: +44 (0)1493 330166

Test Equipment Details: DPI 104 Digital Gauge Serial Number: 3021351 (MDS 07 WTE)
Accuracy Error: +/- 0.05% FSD Date of Last Calibration: 18/10/2012
Calibration Certificate No: 3403-11 Issue By: Antech

INSTRUMENT UNDER TEST

Standard Applied: BS EN 837-1:1998

Manufacturer: Ashford Instruments Serial No: 3227565 (SAD 03 G)
Date of Calibration: 09/04/2013 Ambient Temperature: 8 deg c
Re-Test Date: 09/10/2013 Range: 0 – 230 FHSW
Accuracy: +/- 0.25%

Gauge Reading	Calibration Instrument Reading	Gauge Correction Rising	Gauge Reading	Calibration Instrument Reading	Gauge Correction Falling
0	0	0	0	0	0
20	19.5	-.5	20	19.8	-.2
40	40.2	+.2	40	39.8	-1.2
60	60.2	+.2	60	59.9	-.1
80	80.4	+.4	80	80.2	+.2
100	100.4	+.4	100	100.2	+.2
120	120.4	+.4	120	119.8	-.2
140	140.5	+.5	140	139.9	-.1
160	160.2	+.2	160	160.2	+.2
180	180.2	+.2	180	179.9	-.1
200	199.9	-.1	200	199.7	-.3
220	219.7	-.3	220	219.9	-.1
230	229.6	-.4	230	229.6	-.4

The gauge was lightly tapped before each reading was recorded. The above gauge has been calibrated against equipment traceable to national standards with the results stated and has been found to be satisfactory.

Date: 9-4-13 Signature: 