

CERTIFICATE OF CALIBRATION

Eurocentre, North River Road
Great Yarmouth
Norfolk, NR30 1TE
Tel No: +44 (0)1493 330166

Test Equipment Details: DPI 104 digital gauge Serial Number: MDS07WTE
Accuracy Error: +/- 0.05% FSD Date of Last Calibration: 18th Oct 12
Calibration Certificate No: 2483-12 Issue By: Antech

INSTRUMENT UNDER TEST

Standard Applied: BS EN 837-1:1998

Manufacturer: Wika Serial No: SAD11G
Date of Calibration: 26/02/13 Ambient Temperature: 15 deg c
Re-Test Date: 26/08/13 Range: 0 – 600 psi
Accuracy: +/- 2.5% (15 psi)

Gauge Reading	Calibration Instrument Reading	Gauge Correction Rising	Gauge Reading	Calibration Instrument Reading	Gauge Correction Falling
0	0	0	0	8.3	0
50	53.8	+3.8	50	50.1	+1
100	104.1	+4.1	100	99.7	- .3
200	205.5	+5.5	200	202.8	+2.8
300	305.5	+5.5	300	306.3	+6.3
400	407.0	+7.0	400	408.5	+8.5
500	509.4	+9.4	500	508.1	+8.1
600	614.2	+14.2	600	613.8	+13.8

The gauge was lightly tapped before each reading was recorded. The above gauge has been calibrated against equipment traceable to national standards with the results stated and has been found to be satisfactory.

Date: 13/03/13

Signature: *Antech*