

CERTIFICATE OF CALIBRATION

Eurocentre, North River Road
Great Yarmouth
Norfolk, NR30 1TE
Tel No: +44 (0)1493 330166

Test Equipment Details: DPI 104 digital gauge Serial Number: MDS07WTE
Accuracy Error: +/- 0.05% FSD Date of Last Calibration: 18th Oct 12
Calibration Certificate No: 2483-12 Issue By: Antech

INSTRUMENT UNDER TEST

Standard Applied: BS EN 837-1:1998

Manufacturer: Wika Serial No: SAD12G
Date of Calibration: 26/02/13 Ambient Temperature: 15 deg c
Re-Test Date: 26/08/13 Range: 0 – 600 psi
Accuracy: +/- 2.5% (15 psi)

Gauge Reading	Calibration Instrument Reading	Gauge Correction Rising	Gauge Reading	Calibration Instrument Reading	Gauge Correction Falling
0	0	0	0	7.8	+7.8
50	52.7	+2.7	50	53.4	+3.4
100	102.2	+2.2	100	103.3	+3.3
200	205.1	+5.1	200	204.1	+4.1
300	307.9	+7.9	300	308.4	+8.4
400	407.3	+7.3	400	406.8	+6.8
500	510.0	+10.0	500	510.9	+10.9
600	612.8	+12.8	600	611.4	+11.4

The gauge was lightly tapped before each reading was recorded. The above gauge has been calibrated against equipment traceable to national standards with the results stated and has been found to be satisfactory.

Date: 13/03/13 Signature: *m.anton*