

## CERTIFICATE OF CALIBRATION

Eurocentre, North River Road  
Great Yarmouth  
Norfolk, NR30 1TE  
Tel No: +44 (0)1493 330166

Test Equipment Details: DPI 104 digital gauge      Serial Number: MDS07WTE  
Accuracy Error: +/- 0.05% FSD      Date of Last Calibration: 18<sup>th</sup> Oct 12  
Calibration Certificate No: 2483-12      Issue By: Antech

## INSTRUMENT UNDER TEST

**Standard Applied: BS EN 837-1:1998**

Manufacturer: Wika      Serial No: SAD13G  
Date of Calibration: 26/02/13      Ambient Temperature: 15 deg c  
Re-Test Date: 26/08/13      Range: 0 – 6000 psi  
Accuracy: +/- 2.5% (15 psi)

Gauge Reading	Calibration Instrument Reading	Gauge Correction Rising	Gauge Reading	Calibration Instrument Reading	Gauge Correction Falling
0	0	0	0	95.3	+95.3
500	494.7	-5.3	500	486.9	-13.1
1000	969.7	-30.3	1000	996.4	-3.6
2000	2002.2	+2.2	2000	1999.9	-.1
3000	3002.1	+2.1	3000	2990.9	-9.1
4000	4022.7	+22.7	4000	4036.0	+36
5000	5091.2	+91.2	5000	5096.4	+96.4
6000	6020.0	+20.0	6000	6020.0	+20.0

The gauge was lightly tapped before each reading was recorded. The above gauge has been calibrated against equipment traceable to national standards with the results stated and has been found to be satisfactory.

Date: 13/03/13

Signature: 