

CERTIFICATE OF CALIBRATION

Eurocentre, North River Road
Great Yarmouth
Norfolk, NR30 1TE
Tel No: +44 (0)1493 330166

Test Equipment Details: DPI 104 digital gauge Serial Number: MDS07WTE
Accuracy Error: +/- 0.05% FSD Date of Last Calibration: 18th Oct 12
Calibration Certificate No: 2483-12 Issue By: Antech

INSTRUMENT UNDER TEST

Standard Applied: BS EN 837-1:1998

Manufacturer: Wika Serial No: SAD14G
Date of Calibration: 26/02/13 Ambient Temperature: 15 deg c
Re-Test Date: 26/08/13 Range: 0 – 6000 psi
Accuracy: +/- 2.5% (15 psi)

Gauge Reading	Calibration Instrument Reading	Gauge Correction Rising	Gauge Reading	Calibration Instrument Reading	Gauge Correction Falling
0	0	0	0	35.1	+35.1
500	453.2	-46.80	500	453.7	-46.3
1000	919.1	-80.90	1000	931.6	-68.4
2000	1921.2	-78.80	2000	1917.9	-82.1
3000	2918.5	-81.50	3000	2929.0	-71.00
4000	3947.4	-52.60	4000	3952.7	-47.30
5000	4981.7	-18.30	5000	4992.5	-7.5
6000	5982.1	-17.90	6000	5993.4	-6.6

The gauge was lightly tapped before each reading was recorded. The above gauge has been calibrated against equipment traceable to national standards with the results stated and has been found to be satisfactory.

Date: 13/03/13 Signature: 