

CERTIFICATE OF CALIBRATION

Eurocentre, North River Road
Great Yarmouth
Norfolk, NR30 1TE
Tel No: +44 (0)1493 330166

Test Equipment Details: DPI 104 digital gauge Serial Number: MDS07WTE
Accuracy Error: +/- 0.05% FSD Date of Last Calibration: 18th Oct 12
Calibration Certificate No: 2483-12 Issue By: Antech

INSTRUMENT UNDER TEST

Standard Applied: BS EN 837-1:1998

Manufacturer: Wika Serial No: SAD15G
Date of Calibration: 26/02/13 Ambient Temperature: 15 deg c
Re-Test Date: 26/08/13 Range: 0 – 6000 psi
Accuracy: +/- 2.5% (15 psi)

| Gauge Reading | Calibration Instrument Reading | Gauge Correction Rising | Gauge Reading | Calibration Instrument Reading | Gauge Correction Falling |
|---------------|--------------------------------|-------------------------|---------------|--------------------------------|--------------------------|
| 0 | 0 | 0 | 0 | 43.7 | +43.7 |
| 500 | 421.6 | -78.40 | 500 | 413.2 | -86.80 |
| 1000 | 939.9 | -60.10 | 1000 | 939.7 | -60.30 |
| 2000 | 1983.0 | -17.00 | 2000 | 1991.2 | -8.80 |
| 3000 | 3045.8 | +45.8 | 3000 | 3018.5 | +18.5 |
| 4000 | 4098.4 | +98.4 | 4000 | 4076.8 | +76.8 |
| 5000 | 5112.2 | +112.2 | 5000 | 5126.7 | +126.7 |
| 6000 | 6098.3 | +98.3 | 6000 | 6098.3 | +98.3 |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

The gauge was lightly tapped before each reading was recorded. The above gauge has been calibrated against equipment traceable to national standards with the results stated and has been found to be satisfactory.

Date: 13/03/13 Signature: *Antech*