

## **Analysers, Electrically Powered (battery or mains)**

### **When new**

<b>Examination/Test</b>	<b>Category of Competent Person</b>
In accordance with manufacturer's specification and fit for the purpose it will be used for	3 or 4

### **When in service**

<b>Examination/Test</b>	<b>Validity Period</b>	<b>Category of Competent Person</b>
Examine, confirm working properly and recalibrate to agreed specification in situ	6 months	1, 2 or 4
Note: Analysers are normally calibrated on a much more frequent basis but are not recertified each time	Ongoing	Normally 1 or 2

#### **Note:**

- ◆ Portable analysers can, by definition, be moved around without affecting the validity of their certification.
- ◆ The examination of any analyser should also cover the integrity of any pipework connected to the analyser and providing the sample to it.