



Planned Maintenance Information

Depth/Life Support/Indicating Gauges

SADS ID: SADS/DG/
SADS/SG/
SADS/IG/

IMCA D018 Ref: 18, 19, & 20

Ref	Task No.	Maintenance Guide	Frequency
A	1	General Visual Inspection for damage/corrosion	6 Monthly
	2	Check for accuracy against certified test instrument (DRUCK) – Depth Gauge at 5 points min to 0.25% - Life Support to minimum of 4 points to 2.5%	
	3	Pressurise & Comparison Check Indicating Gauges	
	4	Ensure all gauges function correctly after calibration equipment is removed	
	5	Put new calibration Sticker or 6M visual sticker by gauge	

REMEMBER TO FILL IN RELEVANT PAPERWORK/CERTIFICATION & PASS TO DOCUMENT CONTROLLER FOR SCAN & UPLOAD

Notes :

1. Provision and Use of Work Equipment Regulations 1998 Apply

WARNING: COMPRESSED GAS. During routine maintenance ensure precautions are taken when using HP/LP gases. Wear all relevant PPE. If unsure DO NOT proceed. Seek advice from experienced personnel.